



Phoenix Scale & Food Equipment

Certificate of Calibration

R659 Memorial Highway, Dallas, PA 18612 • Phone (570) 760-3832 • Email phoenixsca@aol.com

Customer: GER Solutions
Address: 795 Roble Road
City, State Zip: Allentown, PENNSYLVANIA 18109


Certificate ID: 140618035171901
ISO Number:
Date: 1/19/2017

Indicator Mfg. Brecknell	Base Mfg. N/A	Cal Date 1/19/2017	Scale ID 140618035	Scale Location Allentown
Indicator Model SB1-505	Base Model N/A	Due Date 1/31/2018	Scale Class III	Scale Range 0 - 5000 lb x 1 lb
Indicator Serial 140618035	Base Serial N/A	Procedure handbook 44	Scale Status In Service	
Test Interval 1 Year				

EQUIPMENT CONDITIONS

[Working] Non-Working Clean Dirty Out of Level Out of Service

SHIFT TEST Shift Test Result: [Pass] Fail Adjust Not Applicable



All tolerances calculated in conformance with Handbook 44 Table 6.

LOAD TEST

Preliminary Load Test		
Test Wt.	Reading	Error
100 lb	100 lb	0 lb
500 lb	500 lb	0 lb
1000 lb	1000 lb	0 lb
2000 lb	2000 lb	0 lb
3000 lb	3000 lb	0 lb
4000 lb	4000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

Final Load Test		
Test Wt.	Reading	Error
100 lb	100 lb	0 lb
500 lb	500 lb	0 lb
1000 lb	1000 lb	0 lb
2000 lb	2000 lb	0 lb
3000 lb	3000 lb	0 lb
4000 lb	4000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

TEST INFORMATION

Test Weight Classification:

E1, F

Traceability Certificate Number(s):

161429 Cal Date: 12/16/2016 Recal Date: 12/31/2019

Standards Used:

howe 50 lb weights 50 lb #5070, 5071, 5072, 5073, 5074, 5075, 5076, 5077, 5078, 5079, 5080

Expanded Uncertainty:

Available on Request or Reported on this Document

Test Location:

[Onsite] Offsite

Overall Result:

[Pass] Fail Adjust

Was the scale within customers required accuracy?

[Yes] No N/A

Environmental Conditions:

[Acceptable] Unacceptable

Comments / Notes:

N/A

Technician:

Rick Shields # 01549